



Molecular Devices i3x Microplate Reader Configuration for Valita®Titer

Parameter	Instrument Settings	
Product	Valita®Titer (#VAL003 / VAL013)	Valita®Titer Plus (#VAL004 / VAL014)
Plate type	96-well: 96 Well Corning Half Area opaque 384 well: 384 Well Corning opaque	
Measurement Mode	Fluorescence Polarization	
Excitation	485 nm	
Emission	535 nm	
Integration time	400 ms	
Read height	4.8 mm	5.2 mm
G-Factor	1.0	
Blank correction	None; 0mg/L standard should be labelled as standard and not blank	